

NOTE: This manual is for three models, which are distinguished by **model A, B** and **C** hereinafter.

A	B	C
		
	Emissivity adjustable	Temperature Alarming

CONTENTS	
1. Safety Information.....	1
2. Precautions.....	1
3. Feature Locations.....	2
4. LCD Illustration.....	4
5. Measurement Principle.....	7
6. Measurement Method.....	7
7. Distance to Spot Ratio.....	9
8. Emissivity.....	9
9. Temperature alarming function.....	13
10. Specifications.....	13
11. Changing the Battery.....	14
12. Accessories.....	15

1. Safety Information

- Please read the following information carefully before using the meter.

- Safety symbols:-

 Danger/Important Information prompt.

 Comply with CE safety standards.

This instrument is compatible with the following standards:

EN61326-1

EN61010-1

EN60825-1

Warning!



Do not point laser directly at eye or indirectly off reflective surfaces.

2. Precautions


- After abrupt ambient temperature changes, allow instrument temperature to stabilize for 30 minutes before using for measurement.

- Avoid operating near strong electromagnetic fields such as arc welders, induction furnaces, etc.
- Do not expose thermometer to excessive ambient temperatures.
- Keep the thermometer clean and avoid getting dust into the detector's optics.
- Do not use solvents to clean the meter.

3. Feature Locations




1. Sighting Groove
2. Laser orifice
3. Detector orifice
4. Tripod nut
5. Trigger
6. Battery cover



7. LCD
8. Laser button
9. Emissivity reduction button (Model B)
Temperature reduction button (Model C)
10. Mode button
11. Emissivity addition button (Model B)
Temperature addition button (Model C)
12. Backlight button
13. °C/°F button


4. LCD illustration

Model A:




1. Laser-On Indicator
2. Holding indicator
3. The first temperature display
4. Low power indicator
5. Average value indicator
6. MAX/MIN indicator
7. The second temperature display
8. °F temperature unit
9. °C temperature unit
10. Emissivity display
11. Measurement indicator

Model B: (Emissivity adjustable)



1. Laser-On Indicator
2. The first temperature display
3. Average value indicator
4. MAX/MIN indicator
5. Emissivity display
6. The second temperature display
7. °F temperature unit
8. °C temperature unit
9. Low power indicator
10. Holding indicator
11. Measurement indicator

Model C: (Temperature alarm)



1. Laser-On Indicator
2. Holding indicator
3. The first temperature display
4. Average value indicator
5. MAX/MIN indicator
6. Temperature alarming indicator
7. The second temperature display
8. °F temperature unit
9. °C temperature unit
10. Low power indicator
11. Emissivity display
12. Measurement indicator

5. Measurement Principle

Infrared thermometers detect infrared energy emitted by objects. The instrument focuses energy through its lens, changing the temperature of a special material that creates an electrical signal. A microcomputer processes this signal and displays the output on the panel of the thermometer. The laser is used only for sighting the instrument and has no other effect on temperature measurement.

6. Measurement Method

1. To measure the temperature of an object, aim the unit at the object and press the trigger. As long as you hold the trigger, you can measure temperature continuously. After releasing the trigger, the LCD will hold the temperature values. The first display indicates the current value. The second display indicates the calculated (Avg., MAX, MIN or MAX-MIN).

2. Use the laser button to activate the laser for aiming at distant objects.

3. Use the backlight button to illuminate the display in low-light conditions.

4. Press the “MODE” button to change the second temperature display. This display can show AVG (average), MAX (maximum), MIN (minimum), MAX-MIN (span) or Alarm (Only model B).


5. Press the °C/°F button to change the unit of measure.

6. (Only model B) Press “+” button to add the emissivity, and press “-” to reduce the emissivity. (Please refer to the chapter 8 “Emissivity”.)


7. (Only model C) Press “MODE” button, when “ALARM” symbol appears, you can set the alarming temperature value. Press “+” to add the alarming temperature value, and press “-” to reduce it. (Please refer to the chapter 9 “Temperature alarming function”.)

7. Distance to Spot Ratio

The thermometer has an optical angle and spot size as shown below.



The target should be larger than the unit's optical spot size. The distance to spot ratio for this thermometer is 12:1 (12mm spot at 1meter) as shown



8. Emissivity

Emissivity is a term used to describe the energy emitting characteristics of a material. The higher the emissivity value a material has, the more

infrared energy it will emit at a particular temperature. And the meter's emissivity is more close to the material's emissivity, the more particular measurement result you can get. So, it is important to set a just emissivity according to the material.

1. Model A and **model C** have a fixed (non-adjustable) emissivity of 0.95. Most organic materials and oxidized metal range in emissivity between 0.85 and 0.98.

2. Model B has an adjustable emissivity from 0.10 to 1.00. For different materials, you can set a different emissivity. An object's emissivity is relative to its material, finish, color and temperature etc. The attached sheet is the emissivity of different substances in the normal temperature.

Substance	Emissivity	Substance	Emissivity
Asphaltum	0.90 - 0.98	Plastic	0.8 – 0.95
Concrete	0.70 – 0.75	Plastic film	0.5 – 0.95
Sand	0.8 – 0.9	Ceramic glaze	0.85 – 0.95
Soil	0.85 – 0.96	Marble	0.90 – 0.94
Water	0.92 – 0.98	Quartz	0.90
Ice	0.96 – 0.98	Aluminium (oxidized)	0.7 – 0.8
Snow	0.83	Copper(oxidized)	0.7 – 0.8
Glass	0.9 – 0.95	Iron (oxidized)	0.78 – 0.82
Keramics	0.9 – 0.94	Lead (oxidized)	0.3
Gypsum	0.89 – 0.91	Zinc (oxidized)	0.1
Red brick	0.75 – 0.90	Zincified iron	0.3
Drygoods	0.93 – 0.98	Cast iron (polished)	0.2
Charcoal	0.96	Iron plate (oxidized)	0.75 – 0.82
Hominine skin	0.98	Steel plate (oxidized)	0.8 – 0.9
Leather	0.75 – 0.80	Stainless steel (Polished)	0.1
Paper	0.8 – 0.94	Gold (polished)	0.1
Lignum	0.8 – 0.9	Silver (polished)	0.1
Asbestos	0.95	Copper(polished)	0.1
Paint	0.9 – 0.98	Aluminium (polished)	0.1
Rubber	0.90 – 0.98	Chromium (polished)	0.1

You also can get an object's emissivity by the following method.

(1) Attach a black adhesive tape or spread some black paint on to the surface.

(2) Set the thermometer's emissivity at 0.95, and then point the meter to the surface of the black adhesive tape (/black paint), and make the measurement. Then you can get a measurement value (say for T1).

(3) Clear the black adhesive tape (/black paint), and point the meter to the same place. Adjust emissivity during measuring; when the measurement value is T1, the emissivity value on LCD is about the object's emissivity.

(4) For some object (e.g. water or with high temperature) that can't be attached a black adhesive tape or spread the black paint, you can use a contact thermometer make measurement first, and then use the same method as the above to get the emissivity.

9. Temperature alarming function

(Only model C)

Press “MODE” button, and when the “ALARM” symbol appears on LCD, you can set the alarming temperature value. Press “+” to add the value, and press “-” to reduce the value. When the measurement temperature is more than the setting value, the inside buzzer will sound.

This function can help operator find the high temperature points by “scan” an object.

10. Specifications

LCD Display: 2-temperature, 4 digits

Distance to Spot Ratio: 12:1

Emissivity:

Fixed at 0.95 (**Model A and C**)

Adjusted from 0.10 to 1.00/step: 0.01 (**Model B**)

Spectral Response: 8-14 μ m

Measurement Range:

-20°C ~ 537°C / -4°F ~ 999°F

Alarming temperature range (Model C):

30°C ~ 535 °C / 86°F ~ 995°F

Accuracy: -20°C ~ 50°C: $\pm 2.5^{\circ}\text{C}$

50°C ~ 537°C $\pm (\text{reading} \times 1\% + 1^{\circ}\text{C})$

Response Time: 1.5 seconds

Laser Power: Less than 1 mW

Auto-off: After 10 seconds

Backlight: White

Operating Environment: 0~ 40°C, 10~90%RH

Storage Environment: -10°C ~ 60°C, <75%RH

Battery: 9Volt, 6F22

Size: 162mm (L) × 56mm (W) × 190mm (H)

11. Changing the Battery

When the battery is nearly exhausted, the low battery symbol will appear on the display. The battery should be replaced soon after this occurs. Pinch the “OPEN” characters to swing the battery cover open.



12. Accessories

- 9V battery.....1 pcs
- Operating manual.....1 pcs
- Carrying bag.....1 pcs